



July 29, 2005

Dr. Donald Cook
Clinical Chief, Laboratory Medicine Program
c/o Lab, St. Clare's

Dear Dr. Cook:

RE: ER/PR Receptors


The following suggestions are made for Dr. Carter's and your best advice on how to proceed from here: *1. Begin recalibration of ER/PR using true positives & negatives*

- 2* **New Testing Results**
 - Send out ER/PR for staining at an outside lab for the next few months. ✓
 - Process the same slides here. ✓
 - Have both sets of slides read by the same Pathologist. — *read in Mt Sinai; read here*
 - Check comparability between stains; and when we are satisfied, revert to processing here.

- 3* **Re-testing**
 - Have the negative results reviewed by pathology type and in consultation with Oncologists to see which cases should be retained and reviewed. ??
 - These cases should follow the same procedure as outlined for #1 above.

4 Ask Vantana to send out one of their technical experts very soon to look at our staining technique.

Yours sincerely,



R.J. WILLIAMS, M.D., M.P.H.
 Vice President, Quality, Diagnostic and
 Medical Services

/dd

General Hospital

Health Sciences Centre, 300 Prince Philip Drive, St. John's, NL, Canada A1B 3V6 Tel. (709) 777-6300 Fax (709) 777-6400
Website: www.hccsj.nf.ca

SITES: Health Sciences Centre (General Hospital/Janeway Children's Health and Rehabilitation Centre/Women's Health Centre)
Dr. Leonard A. Miller Centre • St. Clare's Mercy Hospital • Dr. Walter Tompkins Health Centre • Waterford Hospital