

Completed Staining Run

CIHRT Exhibit P-2309 Page 1

Health Care Corp. of St-Johns, 300 Prince Phillip Drive St-Johns, Newfoundland A1B 3V6

Run Number 1040

Run Operator KEN

Instrument Name Benchmark #1

Run Started 02/26/2004 12:02:17 PM

Instrument Type BenchMark IHC/ISH Staining Module

Run Completed 02/26/2004 3:22:41 PM

Reagent Usage Detail

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
6	BLUING REAGENT	77311	20	9	9	436797	07/06/2005
7	HEMATOXYLIN	109006	20	9	9	432642	07/29/2005
8	I-VIEW H2O2	42381	20	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	20	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	20	1	1	436784 **	04/27/2005
11	I-VIEW INHIBITOR	38091	20	1	1	436784 **	04/27/2005
12	I-VIEW BIOTIN Ig	41943	20	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	20	1	1	436784 **	04/27/2005
14	PROTEASE 1	42308	5	1	1	434324	10/15/2005
15	PREP KIT 8 (CK20 CYTOKERATIN 20)	1627	3	18	18	33102	02/22/2005
16	PREP KIT 7 (CYTO COCKTAIL AE1AE3)	1597	3	17	17	0053A	02/05/2005
17	PREP KIT 6 (HMB45)	1857	5	21	21	1:50	02/16/2005
18	CD3 (PS1)	4037	2	0	0	434410	05/13/2005
19	PREP KIT 4 (CYTO COCKTAIL AE1AE3)	2056	7	0	0	0053A	02/05/2005

Protocol Detail

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
1	CD3 MILD 32	6	_____	+ / -	_____	_____	_____
2	CD3 MILD 32	6	_____	+ / -	_____	_____	_____
3	HMB45 PROT 24 32	***	_____	+ / -	_____	_____	_____
4	HMB45 PROT 24 32	***	_____	+ / -	_____	_____	_____
5	HMB45 PROT 24 32	***	_____	+ / -	_____	_____	_____
6	HMB45 PROT 24 32	***	_____	+ / -	_____	_____	_____
7	HMB45 PROT 24 32	***	_____	+ / -	_____	_____	_____
8	AE1AE3 MILD 32	39	_____	+ / -	_____	_____	_____
9	AE1AE3 MILD 32	39	_____	+ / -	_____	_____	_____
10	AE1AE3 MILD 32	39	_____	+ / -	_____	_____	_____
11	AE1AE3 MILD 32	39	_____	+ / -	_____	_____	_____
12	AE1AE3 MILD 32	39	_____	+ / -	_____	_____	_____
13	AE1AE3 MILD 32	39	_____	+ / -	_____	_____	_____
14	AE1AE3 MILD 32	39	_____	+ / -	_____	_____	_____
15	PR MILD 32	***	_____	+ / -	_____	_____	_____
16	PR MILD 32	***	_____	+ / -	_____	_____	_____
17	PR MILD 32	***	_____	+ / -	_____	_____	_____
18	ER MILD 32	***	_____	+ / -	_____	_____	_____
19	ER MILD 32	***	_____	+ / -	_____	_____	_____
20	ER MILD 32	***	_____	+ / -	_____	_____	_____

* Remaining dispenses are as of time of this report ** Indicates master lots

*** Indicates protocol modified or deleted since run

Completed Staining Run

Health Care Corp. of St-Johns, 300 Prince Philip Drive St-Johns, Newfoundland A1B 3V6

Run Number 1040

Run Operator KEN

Instrument Name Benchmark #1

Run Started 02/26/2004 12:02:17 PM

Instrument Type BenchMark IHC/ISH Staining Module

Run Completed 02/26/2004 3:22:41 PM

Bulk Usage Detail

Bulk Name	Application Slide Count	Lot # / MasterLot #	Expiration Date
CC1	15	435554	06/16/2004
EZ Prep	20	433024	11/08/2004
LCS	20	435946	07/05/2005
Reaction Buffer	20	431298	12/19/2005

System Messages

Message	Notification Time	Sign Off
SSC bottle not full.	02/26/2004 11:59:30 AM	KEN
12-0: Staining module program ran to completion	02/26/2004 3:23:04 PM	LOGINS DISABLED

* Remaining dispenses are as of time of this report ** Indicates master lots

*** Indicates protocol modified or deleted since run

Health Care Corp. of St-Johns, 300 Prince Philip Drive St-Johns, Newfoundland A1B 3V6
NexES v9.20

Printed 07/17/2008 4:42:06 PM

Page 2 of 2

CIHRT Exhibit P-2309 Page 3
Completed Staining Run

Extended Report

Health Care Corp. of St-Johns, 300 Prince Philip Drive St-Johns, Newfoundland A1B 3V6

Run Number 1040

Run Operator KEN

Instrument Name Benchmark #1

Run Started 02/26/2004 12:02:17 PM

Instrument Type BenchMark IHC/ISH Staining Module

Run Completed 02/26/2004 3:22:41 PM

Run Report 1040

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
1	CD3 MILD 32	6		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
18	CD3 (PS1)	4037	Yes	0	0	434410	05/13/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
2	CD3 MILD 32	6		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
18	CD3 (PS1)	4037	Yes	0	0	434410	05/13/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
3	HMB45 PROT 24 32	***		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
14	PROTEASE 1	42308	Yes	1	1	434324	10/15/2005
17	PREP KIT 6 (HMB45)	1857	Yes	21	21	1 :50	02/16/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

* Remaining dispenses are as of time of this report ** Indicates master lots

*** Indicates protocol modified or deleted since run

Run Report 1040

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
4	HMB45 PROT 24 32	***		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
14	PROTEASE 1	42308	Yes	1	1	434324	10/15/2005
17	PREP KIT 6 (HMB45)	1857	Yes	21	21	1 :50	02/16/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
5	HMB45 PROT 24 32	***		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
14	PROTEASE 1	42308	Yes	1	1	434324	10/15/2005
17	PREP KIT 6 (HMB45)	1857	Yes	21	21	1 :50	02/16/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
6	HMB45 PROT 24 32	***		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
14	PROTEASE 1	42308	Yes	1	1	434324	10/15/2005
17	PREP KIT 6 (HMB45)	1857	Yes	21	21	1 :50	02/16/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

* Remaining dispenses are as of time of this report ** Indicates master lots
 *** Indicates protocol modified or deleted since run

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
7	HMB45 PROT 24 32	***		+ / -			
Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
14	PROTEASE 1	42308	Yes	1	1	434324	10/15/2005
17	PREP KIT 6 (HMB45)	1857	Yes	21	21	1 :50	02/16/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
8	AE1AE3 MILD 32	39		+ / -			
Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
19	PREP KIT 4 (CYTO COCKTAIL AE1AE3)	2056	Yes	0	0	0053A	02/05/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
9	AE1AE3 MILD 32	39		+ / -			
Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
19	PREP KIT 4 (CYTO COCKTAIL AE1AE3)	2056	Yes	0	0	0053A	02/05/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

* Remaining dispenses are as of time of this report ** Indicates master lots
 *** Indicates protocol modified or deleted since run

Run Report 1040

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
10	AE1AE3 MILD 32	39		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
19	PREP KIT 4 (CYTO COCKTAIL AE1AE3)	2056	Yes	0	0	0053A	02/05/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
11	AE1AE3 MILD 32	39		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
19	PREP KIT 4 (CYTO COCKTAIL AE1AE3)	2056	Yes	0	0	0053A	02/05/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
12	AE1AE3 MILD 32	39		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
19	PREP KIT 4 (CYTO COCKTAIL AE1AE3)	2056	Yes	0	0	0053A	02/05/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

* Remaining dispenses are as of time of this report ** Indicates master lots
 *** Indicates protocol modified or deleted since run

Run Report 1040

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
13	AE1AE3 MILD 32	39		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot #**	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
19	PREP KIT 4 (CYTO COCKTAIL AE1AE3)	2056	Yes	0	0	0053A	02/05/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
14	AE1AE3 MILD 32	39		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot #**	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
19	PREP KIT 4 (CYTO COCKTAIL AE1AE3)	2056	Yes	0	0	0053A	02/05/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
15	PR MILD 32	***		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot #**	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
16	PREP KIT 7 (CYTO COCKTAIL AE1AE3)	1597	Yes	17	17	0053A	02/05/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

* Remaining dispenses are as of time of this report ** Indicates master lots
 *** Indicates protocol modified or deleted since run

Run Report 1040

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
16	PR MILD 32	***		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
16	PREP KIT 7 (CYTO COCKTAIL AE1AE3)	1597	Yes	17	17	0053A	02/05/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
17	PR MILD 32	***		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
16	PREP KIT 7 (CYTO COCKTAIL AE1AE3)	1597	Yes	17	17	0053A	02/05/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
18	ER MILD 32	***		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
15	PREP KIT 8 (CK20 CYTOKERATIN 20)	1627	Yes	18	18	33102	02/22/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

* Remaining dispenses are as of time of this report ** Indicates master lots
 *** Indicates protocol modified or deleted since run

Run Report 1040

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
19	ER MILD 32	***		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
15	PREP KIT 8 (CK20 CYTOKERATIN 20)	1627	Yes	18	18	33102	02/22/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

Slide Position	Protocol	Protocol #	Case ID+Block ID+Slide ID	Staining	Background	Comments	Sign Off
20	ER MILD 32	***		+ / -			

Reagent Position	Reagent	Serial #	Tests Dispensed	Dispenses Remaining*	Dispenser Life*	Lot # / MasterLot # **	Expiration Date
11	I-VIEW INHIBITOR	38091	Yes	1	1	436784 **	04/27/2005
15	PREP KIT 8 (CK20 CYTOKERATIN 20)	1627	Yes	18	18	33102	02/22/2005
12	I-VIEW BIOTIN Ig	41943	Yes	1	1	436784 **	04/27/2005
13	I-VIEW SA-HRP	34394	Yes	1	1	436784 **	04/27/2005
9	I-VIEW DAB	39339	Yes	1	1	436784 **	04/27/2005
8	I-VIEW H2O2	42381	Yes	1	1	436784 **	04/27/2005
10	I-VIEW COPPER	43962	Yes	1	1	436784 **	04/27/2005
7	HEMATOXYLIN	109006	Yes	9	9	432642	07/29/2005
6	BLUING REAGENT	77311	Yes	9	9	436797	07/06/2005

Bulk Usage Detail

Bulk Name	Application	Slide Count	Lot # / MasterLot #	Expiration Date
CC1		15	435554	06/16/2004
EZ Prep		20	433024	11/08/2004
LCS		20	435946	07/05/2005
Reaction Buffer		20	431298	12/19/2005

System Messages

Message	Notification Time	Sign Off
SSC bottle not full.	02/26/2004 11:59:30 AM	KEN
12-0: Staining module program ran to completion	02/26/2004 3:23:04 PM	LOGINS DISABLED

* Remaining dispenses are as of time of this report ** Indicates master lots
 *** Indicates protocol modified or deleted since run