

July 21/2005 10:30 AM

Dr. Bob Carter
Dr. Jay Cook
Dr. Bill Williams

- i put didn't seem to be a clear picture
- set up case - removed old slides -
- program ~~with~~ ~~and~~ not doing -
new a control
- dew test didn't work

- Dr. Carter feels there was a problem in 2002
- some runs of retests were not saved
- uncertainty for one batch of controls

- current Ventana test is ~~the~~ picking up too much
have set out a guideline of results & sent to
Ms. Susan

- Dr Carter also doing some work in Q.C. at
at use the as controls

- important Dr. Carter have all reports for ER/PR
for each year

- tests may need to be re-trained in immunoprecipitation
at need controlled access to the room
- timing of tests - immunohistochemistry

- need separate lanes, need Q.A.
at proficiency testing

- need to have an external control done by the lab
at Q.A.

Note #7

July 21, 2005 at 10:30 a.m.

**Attendees: Dr. Bev Carter
Dr. Don Cook
Dr. Bob Williams**

- In past didn't seem to be a clear picture.
- Sentinal case – reviewed old slides; program would not always run a control.
- Clear test didn't work.
- Dr. Carter feels there was a problem in 2002.
- Some runs on retrospect were not normal.
- Inconsistency from one batch to another.
- Current Ventana test is picking up too much. Have sent out a _____ results and sent to Mount Sinai.
- Dr. Carter also doing some work on quality control and use them as controls.
- Important for Dr. Carter to have all reports for ER/PR for each year.
- Techs may need to be retrained in immunoperoxidase and need controlled access to the room.
- Training of Techs in immunohistochemistry.
- Need separate service, need Q.A. and proficiency testing.
- Need to have an external consultant to come to the lab and do Q.A.