

July 22, 2005 8:15 PM

Spoke to Dr. Ann O'Brien, director of lab St. Saint John

- Currently have 2 Dako machines, bad method
- ID problem with improper aliquot of Ab to slide, requires constant surveillance with look to ~~see~~ ensure Ab does

Interact with ~~know~~ Dako Netful, sent up representatives who could offer no ~~good~~ explanation, nor resolved the problem.

- County has no state of ER & PR

Positively Rates

- has ID problem with create staining particularly with S-100 protein and a false positive neuropath-immunostain
- County has 1 lab solely devoted to immunos room, separate from histology
- County uses Damm Ab as Mount Sinai in ER & PR.
- uses ASCP checkpath sample a proficiency testing.